

# DIAMOND HARBOUR WOMEN'S UNIVERSITY

## **NOTICE INVITING e-TENDER**

Online Tenders are invited for Delivery and Installation of instruments of Chemistry Laboratory. The works includes Delivery and Installation of Chemistry Laboratory instruments of Diamond Harbour Women's University.

Reputed suppliers/ vendors / bidders having sufficient experience and credentials for successful completion of "Similar Nature" of work in a Government Department/PSU/Autonomous Body/University or any reputed organization. Bidder must have adequate Service Personnel for providing on-site warranty service within the stipulated time.

1	Tender No. & Date	E-NIT/Chemlab/22/DHWU/2019 dated 19.07.2019
2	Brief description of Job	Delivery and Installation of Chemistry Laboratory instruments of Diamond Harbour Women's University.
3	Tender Fee	NIL
4	Earnest Money Deposit	Rs.30,000.00 (Rupees Thirty Thousand only) in the form of Demand Draft from any Scheduled bank in favour of Diamond Harbour Women's University at Sarisha
5	Date of Downloading / Sale of Tender 19.07.2019 document	19.07.2019 from 5.00 pm
6	Bid Submission Start date & time	19.07.2019 from 5.00 pm
7	Last date & time of EMD & Tender Fee submission (offline)	03.08.2019 at 12.30 pm
8	Last date & time of Bid Submission	03.08.2019 at 12.30 pm
9	Date & time of Technical Bid Opening	05.08.2019 at 12.30 pm
10	Submission of EMD	DIAMOND HARBOUR WOMEN'S UNIVERSITY, Sarisha, Diamond Harbour Road, 24 Pgs (S), 743368
11	Contact	(03174) 245801 / 802

1. Intending bidder may download the tender documents from the website <https://wbtenders.gov.in> directly with the help of Digital Signature Certificate. Necessary tender fee may be remitted through Demand Draft issued from any Scheduled Bank in favour of "DIAMOND HARBOUR WOMEN'S UNIVERSITY", payable at Sarisha and also to be documented through e-filing. Cost of Earnest Money Deposit (EMD) may be remitted through Demand Draft issued from any Scheduled Bank in favour of "DIAMOND HARBOUR WOMEN'S UNIVERSITY", payable at Sarisha and also to be documented through e-filing. The original Demand Draft against Tender Fees & Earnest Money Deposit (EMD) should be submitted physically to the office of DIAMOND HARBOUR WOMEN'S UNIVERSITY, Sarisha, Diamond Harbour Road, 24 Pgs (S), 743368 under sealed cover on or before 11:00 Hrs of 03.08.2019. In this E-NIT tender fee is not required but EMD is required.
2. Both Techno Commercial Bid and Financial Bid are to be submitted concurrently duly digitally signed in the website <https://wbtenders.gov.in>
3. Tender documents may be downloaded from website and submission of Techno Commercial Bid and Financial Bid will be done as per Time Schedule stated in Section – C of this Tender Document.
4. The Financial Bid of the prospective Bidder will be considered only if the Techno Commercial Bid of the bidder is found qualified by the Tender Committee. The decision of the 'Tender Committee' will be final and absolute in this respect. The list of Qualified Bidders will be displayed in the website.

## **CONTENTS OF THE TENDER DOCUMENT**

The Tender document comprises of the following:

SECTION – A	SCOPE OF WORK & RESPONSIBILITY
SECTION – B	ELIGIBILITY CRITERIA
SECTION – C	DATE AND TIME SCHEDULE
SECTION – D	INSTRUCTION TO BIDDER
SECTION – E	BID FORM
SECTION – F	TECHNO COMMERCIAL EVALUATION & AWARDING OF CONTRACT
SECTION – G	GUIDANCE FOR E-TENDERING
SECTION – H	BILL OF MATERIAL
SECTION – I	TECHNICAL SPECIFICATION WITH COMPLIANCE STATEMENT
SECTION – J	DETAILS OF ORDERS EXECUTED BY BIDDER FINANCIAL
SECTION – K	CAPABILITY OF BIDDER
SECTION – L	BIDDER’S DETAILS
SECTION – M	MANUFACTURER’S AUTHORIZATION FORM
SECTION – N	PRE-BID MEETING QUERY
SECTION – O	LIST OF CLIENTS OF SIMILAR ORDERS
SECTION – P	PROFORMA FOR PERFORMANCE BANK
SECTION – Q	GUARANTEE NIT DECLARATION

## **SECTION – A**

### **SCOPE OF WORK & RESPONSIBILITY**

#### **For vendors / bidders:**

- 1) **Supply:** Entrusted bidder must supply and work as per work order. All supplies must be certified. Supplied components must satisfy the specifications and guidelines, as applicable. Bidder is allowed to propose only one make / model per system or software or items required for the proposed solution. Bids with optional offers would be summarily rejected.
- 2) **Installation, Integration & Implementation:** Delivery & Installation will be governed by the stated guidelines and associated standards & in line with the scope of work defined in this document. The delivery / supply of the materials done by successful Bidder at respective location only. The Post installation warranty supports are the sole responsibility of the Original Equipment Manufacture (OEM) and /or bidder.
- 3) **Service Quality and Assurance:** Services within installation, warranty period must be rendered with OEM for hardware.
- 4) **Warranty:** All hardware component service items should have minimum warranty period of period i.e. 12 months from the date of final acceptance of the system by the user. Within warranty period, entrusted OEM will replace, re-integrate faulty, buggy components.
- 5) **Delivery:** All materials are to be delivered at office of the Diamond Harbour Women's University. After necessary checking, testing, marking of the all materials under supervision of DHWU. Successful bidder will take necessary step to transport the material at installation site and take up the installation there at their own cost.
- 6) **Warranty Certificate:** On successful installation bidder to get installed equipments with actual serial no. duly certified by respective OEM for minimum one year warranty coverage. Any change in equipment serial no. should be reflected in OEM website or OEM declaration in their Letter head. Final Bills will be submitted after completion of this activity.

## SECTION – B

### ELIGIBILITY CRITERIA

1. The bidder must be a company registered under Companies Act, 1956/2013 or Partnership or LLP or OPC or Proprietary Firm. Documentary (Certificate of incorporation/Relevant document) evidence to be submitted.
2. The bidder should have their presence in Kolkata with own office. Valid proof should be submitted along with the bid.
3. The bidder should have valid GST Registration Certificate, PAN & Trade License. Bidder shall have to submit photocopy of the documents.
4. The bidder shall have executed “Similar Nature” of single order.
5. Bidder should submit Earnest Money Deposit (EMD) of Rs. 30000.00 (Rupees Thirty Thousand Only) in the form of Demand Draft from any Scheduled Bank in favour of DIAMOND HARBOUR WOMEN’S UNIVERSITY payable at Sarisha.
6. Manufacturers tender specific authorizations for Chemistry Laboratory Instruments must be submitted as per format enclosed (Section - M).
7. The bidder shall submit Bid Form (Section – E) duly signed by the authorized signatory of the company as per the format enclosed. Deviation in format may not be accepted.
8. The bidder shall not have been blacklisted by any University/State/Central Government or PSU Organization or bilateral/multilateral funding agencies for breach of ethical conduct or fraudulent practices as on date of submission of the proposal. **Declaration on bidder’s letter head to be submitted.**

**SECTION – C**

**DATE AND TIME SCHEDULE**

<b>Sl. No.</b>	<b>Particulars</b>	<b>Date &amp; Time</b>
1	Date of uploading of N.I.T. & other Documents (online) (Publishing Date)	19.07.2019 from 5.00 pm
2	Documents download / sale start date (Online)	19.07.2019 from 5.00 pm
3	Last Date and time of sending the queries (Offline)	30.07.2019 from 12.30 pm
4	Corrigendum, if any will be published (On Line)	NA
5	Bid Submission start date & time (On line)	19.07.2019 from 5.00 pm
6	Last Date & time of submission of original Demand Draft/Pay Order for cost of Earnest Money Deposit (Off line)	03.08.2019 from 12.30 pm
7	Bid Submission closing date & time (On line)	03.08.2019 from 12.30 pm
8	Bid opening date & time for Technical Proposals (Online)	05.08.2019 from 12.30 pm
9	Date for opening of Financial Bid (Online)	To be notified later

## SECTION – D

### INSTRUCTION TO BIDDER

#### 1. DEFINITIONS

In this document, the following terms shall have following respective meanings:

“Acceptance Test Document” means a document, which defines procedures for testing the functioning of installed system. The document will be finalized with the contractor within 7 days of issuance of the Letter of Award.

“Bidder” means any firm offering the solution(s), service(s) and /or materials required in the tender. The word Bidder when used in the pre award period shall be synonymous with Bidder, and when used after award of the Contract shall mean the successful Bidder.

“Contract” is used synonymously with Agreement.

“Contract Price” means the price to be paid to the Contractor for providing the Solution, in accordance with the payment terms.

“Contractor” means the Bidder whose bid to perform the Contract has been accepted by Tender Committee and is named as such in the Letter of Award.

“Default Notice” mean the written notice of Default of the Agreement issued by one Party to the other.

“Installation” means installation of Instruments of Chemistry Laboratory of Diamond Harbour Women’s University.

“Fraudulent Practice” means a misrepresentation of facts in order to influence a procurement process or the execution of a Contract and includes collusive practice among Bidders (prior to or after Bid submission) designed to establish Bid prices at artificial noncompetitive levels.

“Good Industry Practice” shall mean the exercise of that degree of skill, diligence and prudence which would reasonably and ordinarily be expected from a reasonably skilled and experienced Operator engaged in the same type of undertaking under the same or similar circumstances.

“GoI” shall stand for the Government of India.

“GoWB” means Government of West Bengal

“Personnel” means persons hired by the Bidder as employees and assigned to the performance of the Infrastructure Solution or any part thereof.

“Similar Nature of Work” means of Delivery and Installation of Chemistry Laboratory instruments of Diamond Harbour Women’s University.

“Project” Delivery and Installation of Chemistry Laboratory instruments of Diamond Harbour Women’s University.

“Services” means the work to be performed by the Bidder pursuant to this Contract, as described in the detailed Scope of Work.

“Interest rate” means “364 days Government of India (GoI) Treasury Bills” rate.

"Law" shall mean any Act, notification, bye law, rules and regulations, directive, ordinance, order or instruction having the force of law enacted or issued by the Central Government and/or the Government of West Bengal or any other Government or regulatory authority or political subdivision of government agency.

“LOI” means issuing of Letter of Intent shall constitute the intention of the DHWU to place the Purchase Order with the successful bidder.

"Operator" means the company providing the services under Agreement

“Requirements” shall mean and include schedules, details, description, statement of technical data, performance characteristics, standards (Indian as well as International) as applicable and specified in the Contract.

“Service” means provision of Contracted service viz., operation, maintenance and associated services for DEPLOYED SYSTEMS as per Section titled “Scope of Work”

“Termination Notice” means the written notice of termination of the Agreement issued by DHWU.

"Uptime" means the time period when specified services are available with specified technical and service standards as mentioned in section titled WARRANTY SUPPORT” "%Uptime" means ratio of 'up time' (in minutes) as mentioned in section titled “Warranty support”

"Service Down Time" (SDT) means the time period when specified services with specified technical and operational requirements as mentioned in section titled "WARRANTY SUPPORT" are not available to DHWU and its user departments and organizations.

"DHWU" means DIAMOND HARBOUR WOMEN'S UNIVERSITY.

## **2. COST OF BIDDING**

The bidder shall bear all costs associated with the preparation and submission of the bid and DHWU will in no case be responsible for those costs regardless of the conduct or outcome of the bidding process.

## **3. BID DOCUMENT**

Bidder is expected to examine all instructions, forms, terms and requirements in the bid document. The invitation to bid together with all its attachments thereto shall be considered to be read, understood and accepted by the bidder unless deviations are specifically stated in the invitation by the bidder. Failure to furnish all information required by the bid document or a bid not substantially responsive to the bid document in every respect may result in the bid being rejected.

## **4. AMENDMENT OF BID DOCUMENT**

At any time prior to the deadline for submission of proposals, DHWU reserves the right to add/modify/delete any portion of this document by issuance of a Corrigendum, which would be published on the website and will also be made available to all the Bidders who have been issued the tender document. The Corrigendum shall be binding on all bidders and will form part of the bid documents.

## **5. MODIFICATION AND WITHDRAWAL OF BIDS**

As per the bidding process available in the tender.

## **6. LANGUAGE OF BID & CORRESPONDENCE**

The proposal will be prepared by the Bidder in English language only. All the documents relating to the proposal (including brochures) supplied by the firm should also be in English, and the correspondence between the Bidder & DHWU will be in English language only. The correspondence by fax/E-mail must be subsequently confirmed by a duly signed formal copy.

## **7. BIDDER'S SOLUTION**

The bidders are requested to study the Bill of Materials supplied with this document carefully. While working out the solution the bidder has to work with the broad specification provided in the tender documents, conforming to the model, make and Part number (wherever provided). While submitting the bid the bidder has to detail out all components needed to complete the system BOM. The bidder is required to quote for each item retaining all major components/sub system detailed and specified. As the contractor will be responsible for smooth functioning of the system, availability of spares during the tenure of the warranty period have to be taken care of by the contractor to maintain the guaranteed uptime.

## **8. EARNEST MONEY DEPOSIT (EMD)**

The firm shall furnish an EMD of Rs. 30,000.00/- (Rupees Thirty Thousand Only) in the form of a demand draft from a Scheduled Bank payable at Sarisha and in favour of DIAMOND HARBOUR WOMEN'S UNIVERSITY. Any bid not accompanied with the EMD shall be rejected.

## **9. FORFEITURE OF EMD**

EMD made by Bidder may be forfeited under the following conditions: If Bidder withdraws the proposal before the expiry of validity period.

During the evaluation process, if a Bidder indulges in any such activity as would jeopardize the process, the decision of DHWU regarding forfeiture of EMD shall be final and shall not be called upon question under any circumstances.

If Bidder violates any of the provisions of the terms and conditions of the proposal. In the case of a successful Bidder, if Bidder fails to:

- a) Accept the work order along with the terms and conditions.
- b) Furnish performance security.
- c) Violates any of the work conditions of this proposal or indulges in any such activities as would jeopardize the work.
- d) Submitting false/misleading information/declaration/documents/proof/etc.

The decision of DHWU regarding forfeiture of EMD shall be final and shall not be called upon to question under any circumstances, besides, forfeiture of EMD even the Bidder will be deferred from participating in any job for a period of one year.

## **10. FORMS AND FORMATS**

The various inputs for the Techno Commercial and Financial Bids are to be submitted in the format specified. The bidder shall use the form, wherever specified, to provide relevant information. If form does not provide space for any required information, space at the end of the form or additional sheets shall be used to convey the said information. For all other cases, the bidder shall design a form to hold the required information.

## **11. LACK OF INFORMATION TO BIDDER**

The bidder shall be deemed to have carefully examined the Bid document to his entire satisfaction. Any lack of information shall not relieve the bidder of his responsibility to fulfill his obligation under the bid.

## **12. CONTRACT EXECUTION**

On receipt of the Letter of Award/Purchase Order the contractor should submit a Performance Bank Guarantee (PBG) equivalent to 10% of the total contract value within three weeks from the date of receipt of Letter of Award/Purchase Order. The PBG should be valid for six months more than the warranty period. All delivery of the material will have to be completed within 30 days from the date of acceptance of contract and the contractor has to ensure all activities leading to the commissioning of the contract to be completed within 45 days from the date of award. Subsequent to the award of contract, the contractor will have to arrange for the requisite material as per BOM.

## **13. TIME SCHEDULE FOR DELIVERY & INSTALLATION**

All delivery & Installation of the material will have to be completed within 4-6 weeks from the date of Order.

## **15. LOCATION OF DELIVER & INSTALLATION**

The office of Diamond Harbour Women's University, Sarisha, Diamond Harbour Road, 24 Pgs (S), 743368

## **16. LIQUIDATED DAMAGE / PENALTY**

The job includes the supply and installation of materials mentioned in the tender document. In the event of failure to meet the job completion in stipulated date/time liquidated damage may be imposed on the contractor for sum not less than 0.5% of the contract value for that item/job for each week or part thereof, subject to a ceiling of 10% of the total contract value (excluding all taxes & duties and other charges). In the event of LD exceeds 10% of the order value, DHWU reserves the right to terminate the contract and DHWU will get the job completed by any other competent party. The difference of cost incurred by DHWU will be recovered from the contractor and PBG will be invoked.

## **17. LIABILITY**

In case of a default on bidder's part or other liability, DHWU shall be entitled to recover damages from the Contractor. In each such instance, regardless of the basis on which DHWU is entitled to claim damages from the Contractor (including fundamental breach, negligence, misrepresentation, or other contract or tort claim), Contractor shall be liable for no more than:

- Payment referred to in the Patents and Copyrights clause.
- Liability for bodily injury (including death) or damage to real property and tangible personal property limited to that cause by the Contractor's negligence.
- As to any other actual damage arising in any situation involving non-performance by Contractor pursuant to or in any way related to the subject of this Agreement, the charge paid by DHWU for the individual product or Service that is the subject of the Claim. However, the contractor shall not be liable for
- For any indirect, consequential loss or damage, lost profits, third party loss or damage to property or loss of or damage to data.

For any direct loss or damage that exceeds the total payment for Contract Price made or expected to be made to the Contractor hereunder.

## **18. PATENTS & COPYRIGHT**

If a third party claims that a product delivered by the Contractor to DHWU infringes that party's patent or copyright, the Contractor shall defend DHWU against that claim at Contractor's expense and pay all costs, damages, and attorney's fees that a court finally awards or that are included in a settlement approved by the Contractor, provided that DHWU.

- Promptly notifies Contractor in writing of the claim
- Allows Contractor to control and co-operate with Contractor in the defense and any related settlement negotiations.

Remedies: If such a claim is made or appears likely to be made, DHWU would permit Contractor to enable DHWU to continue to use the product, or to modify it, or replace it with one that is at least functionally equivalent. If Contractor determines that none of these alternatives is reasonably available, DHWU agrees to return the product to Contractor on Contractor's written request. Contractor will then give DHWU a credit equal to for a machine. DHWU's net book value (provided DHWU has followed generally accepted accounting principles for a generally available software product produced by Contractor (Program) the amount paid by DHWU or 12 months charges (which ever is lesser) and for materials the amount paid by DHWU for the materials. These will be Contractor's entire obligation regarding any claim of infringement.

## **19. SUSPENSION OF WORK**

DHWU shall have the power at any time and from time to time by notice to the Contractor to delay or suspend the progress of the work or any part of the work due to any other adequate reasons and on receipt of such notice the contractor shall forthwith suspend further progress of the work until further notice from DHWU. The Contractor shall recommence work immediately after receiving a notice to do so from DHWU. The whole or any part of the time lost for such delay or suspension shall, if DHWU in its absolute discretion thinks fit, but not otherwise, be added to the time allowed for completion.

## **20. TERMS OF PAYMENT**

90% payment will be made after satisfactory delivery and installation of the material duly certified by the HOD/P.I and remaining 10% will be made after completion of warranty period.

## **21. GOVERNING LAWS**

This contract should be governed by and interpreted by Arbitration clause in accordance with Laws in force in India. The courts at Diamond Harbour/Kolkata shall have exclusive jurisdiction in all matters arising under the contract. The selected vendor shall keep himself fully informed of all current national, state and municipal law and ordinances. The selected vendor shall at their own expense, obtain all necessary permits and license and pay all fees and taxes required by law. These will be selected vendor's entire obligation regarding any claim of infringement. The selected vendor hereto agrees that it shall comply with all applicable union, state and local laws, ordinances, regulations and codes in performing its obligations hereunder, including the procurement of licenses, permits certificates and payment of taxes where required. The selected vendor shall establish and maintain all proper records (particularly, but without limitation, accounting



records) required by any law, code/practice of corporate policy applicable to it from time to time including records and returns as applicable under labor legislation.

## **22. CORRUPT OR FRAUDULENT**

The Tender Committee requires that the bidders under this Tender observe the highest standards of ethics during the procurement and execution of such contracts. For this purpose the definition of corrupt and fraudulent practices will follow the provisions of the relevant laws in force. The Tender Committee will reject a proposal for award if it detects that the bidder has engaged in corrupt or fraudulent practices in competing for the contract in question. The Tender Committee will declare a firm ineligible, either indefinitely or for a stated period of time, if it at any time determines that the firm has engaged in corrupt and fraudulent practices in competing for, or in executing, a contract.

## **23. BIDDING CLAUSE**

All decisions taken by the Tender Committee regarding the processing of this tender and award of contract shall be final and binding on all parties concerned.

The Tender Committee reserves the right:

- To vary, modify, revise, amend or change any of the terms and conditions mentioned above and,
- To reject any or all the Tender/s without assigning any reason whatsoever thereof or to annul the bidding process and reject all bids at any time prior to award of contract, without thereby incurring any liability to the affected bidder(s) or any obligation to inform the affected bidder(s) of the grounds for such decision.

## **24. WORKMEN'S COMPENSATION**

In every case in which by virtue of the provision of the workmen's compensation Act 1923 or any other relevant acts and rules, compensation to a workmen employed by the contractor, is payable, then this should be done by the Contractor. If DHWU is obliged to make any compensation under the said rules and acts, then the amount shall be recovered without prejudice, from the bills and due of the Contractor. DHWU shall not be bound to contest any claim made against the Contractor in respect of workmen's compensation.

## **25. CONTRACTOR'S EMPLOYEES**

The Contractor shall comply with the provision of all labour legislation including the requirement of the payment of Wage Act 1936 and the rules framed there under and modifications thereof in respect of men employed by him in carrying out the contract. The Contractor must ensure that he complies with PF, ESI regulation for all his deployed employees. The Contractor shall see that all authorized Sub Contractors under him similarly complied with the above requirement.

## **26. SAFETY MEASURES**

The Contractor shall in the course of execution of the work take all necessary precaution for the protection of all persons and property. The Contractor shall take adequate measures to protect the work and prevent accident during the work. In the event of any accident to any person or persons or damage or injury of any description to any person or property due to failure on the part of the contractor in taking proper precautionary measures the contractor shall be responsible for and must make good the loss the damage at his own cost to the satisfaction of University and employees of the University shall be indemnified from all claims or liabilities arising there from or any expenses incurred on account thereof.

## **27. EQUIPMENT**

All tools & tackles necessary for the work shall have to be procured by the contractor unless otherwise specified elsewhere in these tender documents. The equipment used by the contractor for a particular work must be appropriate for the type of work. The contractor shall maintain the equipment used on the work properly so that they are in good working condition. In no case shall the contractor use defective or imperfect equipment in the work. The contractor shall arrange to replace or repair all defective equipment so that the progress of the work is not hampered. No defective equipment should be left at the site of work and the University shall not be responsible for any loss or damage to any of these equipments during the course of the execution of the work.

## **28. SUB-CONTRACT**

The purchaser (DHWU) does not recognize the existence of Sub-Contractors. The Contractor's responsibility is not transferable.

## **29. TERMINATION FOR DEFAULT**

DHWU may without prejudice to any other remedy or right of claim for breach of contract by giving not less than 30 days written notice of default sent to the contractor, terminate the order in whole or in part. If the contractor materially fails to render any or all the services within the time period specified in the contract or any extension thereof granted by DHWU in writing and fails to remedy its failure within a period of thirty days after receipt of default notice from DHWU. If the project (delivery, commissioning as well as warranty maintenance support is not carried out according to specification due to deficiency in service as per terms of the contract. In such case DHWU will invoke the amount held back from the contractor as PBG.

## **30. BANKRUPTCY**

If the contractor becomes bankrupt or have a receiving order made against him or compound with his creditors or being a corporation commence to be wound up, not being a voluntary winding up for the purpose only or amalgamation or reconstruction, or carry on their business under a receiver for the benefit of their creditors or any of them, DHWU shall be at liberty to terminate the engagement forthwith without any notice in writing to the contractor or to the liquidator or receiver or to any person in whom the contractor may become vested and without any compensation to give such liquidator or receiver or other person the option of carrying out the engagement subject to their providing a guarantee for the due and faithful performance of the engagement up to an amount to be determined by DHWU.

## **31. FORCE MAJEURE**

It is hereby defined as any cause, which is beyond the control of the Contractor or DHWU as the case may be, which such party could not foresee or with a reasonable amount of diligence could not have foreseen and which substantially affect the performance of the contract, such as

- War, Hostilities or warlike operations (whether a state of war be declared or not), invasion, act of foreign enemy and civil war.
- Rebellion, revolution, insurrection, mutiny, usurpation of civil or military, government, conspiracy, riot, civil commotion and terrorist area.
- Confiscation, nationalization, mobilization, commandeering or requisition by or under the order of any government or de facto authority or ruler, or any other act or failure to act of any local state or national government authority.
- Strike, sabotage, lockout, embargo, import restriction, port congestion, lack of usual means of public transportation and communication, industrial dispute, shipwreck, shortage of power supply epidemics, quarantine and plague.
- Earthquake, landslide, volcanic activity, fire flood or inundation, tidal wave, typhoon or cyclone, hurricane, nuclear and pressure waves or other natural or physical disaster.

If either party is prevented, hindered or delayed from or in performing any of its obligations under the Contract by an event of Force Majeure, then it shall notify the other in writing of the occurrence of such event and the circumstances of the event of Force Majeure within fourteen days after the occurrence of such event. The party who has given such notice shall be excused from the performance or punctual performance of its obligations under the Contract for so long as the relevant event of Force Majeure continues and to the extent that such party's performance is prevented, hindered or delayed.

The party or parties affected by the event of Force Majeure shall use reasonable efforts to mitigate the effect of the event of Force Majeure upto its or their performance of the Contract and to fulfill its or their obligation under the Contract but without prejudice to either party's right to terminate the Contract.

No delay or nonperformance by either party to this Contract caused by the occurrence of any event of Force Majeure shall.

- Constitute a default or breach of the contract.
- Give rise to any claim fro damages or additional cost or expense occurred by the delay or nonperformance. If, and to the extent, that such delay or nonperformance is caused by the occurrence of an event of Force Majeure.

### **32. INSURANCE COVERAGE**

Appropriate insurance to cover all solution components for the transit period and until the time of its acceptance at the respective site is to be taken by the contractor. As the contractor will carry the risk for the material in his books during transit, the contractor should arrange insurance for the total system as period from the dispatch till Acceptance Test is successfully achieved. Further the contractor is to take all required insurance coverage in respect of all its personnel who shall be working on this engagement.

### **33. WARRANTY**

The OEM on behalf of bidder will warranty that products supplied under the contract are newly made and are free from defects in the design, engineering and workmanship. The Contractor would be responsible for the up keep and maintenance of the necessary deliverables under the scope of work during the entire warranty period. Within warranty period, entrusted OEM will replace, re-integrate faulty, buggy components.. The Contractor/OEM shall not, without the express prior written consent of DHWU, assign to any third party of the contract or part thereof. Service support for the entire warranty period will be on site and comprehensive (including spares) and free of cost for the entire warranty period. Warranty will be invalid if the equipment is serviced by unauthorized personnel of misuse is detected.

### **34. WARRANTY SUPPORT**

The total system will be warranted against bad workmanship and manufacturing defects from the date of acceptance of the system whole or part. Service support for the entire warranty period will be on site.

### **35. PERFORMANCE BANK GUARANTEE (PBG)**

As a guarantee for timely delivery, installation and commissioning of equipment as well as performance of on-site warranty support, as mentioned in Bill of Material, from the date of final acceptance of systems/items and pertaining to proper running of the systems/items, the bidder will have to submit 10% of the contract value as security in the form of Performance Bank Guarantee from any nationalized bank as per format enclosed (Section – P).

### **36. CONTRACTOR'S RESPONSIBILITIES**

Refer Section – A (Scope of Work & Responsibility)

### **37. NO WAIVER OF RIGHTS**

Neither the inspection by DHWU or any of their agents nor any order by DHWU for payment of money or any payment for or acceptance of the whole or any part of the works by DHWU, nor any extension of time, nor any possession taken by DHWU shall operate as a waiver of any provision of the contract or of any power reserved to DHWU, or any right to damages here in provided, nor shall any waiver of any breach in the contract be held to be a waiver of any other subsequent breach.

### **38. GRAFTS, COMMISSIONS, GIFTS, ETC.**

It is the Purchaser's policy to require that bidders, suppliers, contractors and consultants under contracts, observe the highest standard of ethics during the procurement and execution of such contracts. Any graft, commission, gift or advantage given, promised or offered by or on behalf of the contractor or his partner, agent, officers, director, employee or servant or any one on his or their behalf in relation to the obtaining or to the execution of this or any other contract with DHWU shall in addition to any criminal liability which it may incur, subject the contractor to the cancellation of this and all other contracts and also to payment of any loss or damage to DHWU resulting from any cancellation. DHWU shall then be entitled to deduct the amount so payable from any monies otherwise due to the contractor under contract.

### **39. ENFORCEMENT OF TERMS**

The failure of either party to enforce at any time any of the provision of this contract or any rights in respect thereto or to exercise any option here in provided shall in no way be construed to be a waiver to such provisions, rights or options or in any way to affect the validity of the contract. The exercise by either party of any of its rights herein shall not preclude or prejudice either party from exercising the same or any other right it may have hereunder.

### **40. PERIOD OF VALIDITY OF OFFER**

For the purpose of placing the order, the proposals shall remain valid till 180 days. During the period of validity of proposals, the rates quoted shall not change. In exceptional circumstances, DHWU may ask for extension of the period of validity and such a request shall be binding on Bidders. DHWU's request and the response to such a request by various Bidders shall be in writing. A Bidder agreeing to such an extension will not be permitted to increase its rates.

### **41. TAXES & DUTIES**

- The prices shall be inclusive of all taxes & levies including GST and other statutory duties as applicable. Rate of taxes should be indicated separately in the Price Bid.
- Contract Price specified in Price Bid should be based on the taxes & duties and charges prevailing at the date one day prior to the last date of Bid submission.
- Bidder submitting a bid shall produce valid statutory documents / certificates with respect to GST, Income Tax, ROC, Prof. Tax, Trade Licence, etc. All such documents / certificates shall remain valid on the last date of tender submission.
- GST component of the invoice of the bidder may be kept on hold in case there is any mismatch / irregularity in GST return filling on the part of the bidder.

### **42. DISCREPANCIES IN BID**

- Discrepancy between description in words and figures, the rate which corresponds to the words quoted by the bidder shall be taken as correct.
- Discrepancy in the amount quoted by the bidder due to calculation mistake of the unit rate then the unit rate shall be regarded as firm.
- Discrepancy in totaling or carry forward in the amount quoted by the bidder shall be corrected.

### **43. BID DUE DATE**

The online tender has to be submitted not later than the due date and time specified in the Important Dates Sheet. DHWU may as its discretion on giving reasonable notice by fax, or any other written communication to all prospective bidders who have been issued the bid documents, extend the bid due date, in which case all rights and obligations of the DHWU and the bidders, previously subject to the bid due date, shall thereafter be subject to the new bid due date as extended.

### **44. LATE BID**

Any proposal received by DHWU after the deadline for submission of proposals may not be accepted.

### **45. OPENING OF BID BY DHWU**

Bids shall be opened and downloaded electronically through operation of the process in the e-Tender portal in presence of Tender Committee. Bidders interested to remain present during electronic bid opening may attend the bid opening session at DHWU premises at scheduled date & time.

### **46. CONTACTING DHWU**

Bidder shall not approach DHWU officers beyond office hours and/or outside DHWU office premises from the time of the Bid opening to the time of finalization of successful bidder. Any effort by bidder to influence DHWU office in the decision on Bid evaluation, Bid comparison or finalization may result in rejection of the Bidder's offer. If the bidder wishes to bring additional information to the notice of DHWU, it should be in writing following the procedure mentioned hereinabove.

### **47. DHWU'S RIGHT TO REJECT ANY OR ALL BIDS**

DHWU reserves the right to reject any bid and to annul the bidding process and reject all bids at any time prior to award of Contract, without thereby incurring any liability to the affected bidder(s) or any obligation to inform the affected bidder(s) of the grounds for such decision.

### **48. BID CURRENCIES**

Prices shall be quoted in Indian Rupees, inclusive of all prevailing taxes, levies, duties, cess etc.

### **49. PRICE**

- Price should be quoted in the BOQ format only. No deviation is acceptable.
- Price quoted should be firm, inclusive of packing, forwarding, insurance and freight charges.
- Percentage/specified amount of taxes & duties should be clearly mentioned otherwise DHWU reserves the right to reject such vague offer.
- Price to be quoted inclusive of supply, installation & commissioning charges.

### **50. CANVASSING**

Canvassing or support in any form for the acceptance of any tender is strictly prohibited. Any bidder doing so will render him liable to penalties, which may include removal of this name from the register of approved Contractors.

### **51. NON-TRANSFERABILITY OF TENDER**

This tender document is not transferable.

## **52. FORMATS AND SIGNING OF BID**

The original and all copies of the proposals shall be neatly typed and shall be signed by an authorized signatory (ies) on behalf of the Bidder. The authorization shall be provided by written Power of Attorney accompanying the proposal. All pages of the proposal, except for un-amended printed literature, shall be initialed by the person or persons signing the proposal. The proposal shall contain no interlineations, erase or overwriting. In order to correct errors made by the Bidder, all corrections shall be done & initialed with date by the authorized signatory after striking out the original words/figures completely.

## **53. WITHDRAWAL OF BID**

Bid cannot be withdrawn during the interval between their submission and expiry of Bid's validity period. Fresh Bid may be called from eligible bidders for any additional item(s) of work not mentioned herein, if so required.

## **54. INTERPRETATION OF DOCUMENTS**

If any bidder should find discrepancies or omission in the specifications or other tender documents, or if he should be in doubt as to the true meaning of any part thereof, he shall make a written request to the tender inviting authority for correction/clarification or interpretation or can put in a separate sheet along with his technical bid document.

## **55. PREPARATION OF TENDER**

Tender shall be submitted in accordance with the following instructions:

- a) Tenders shall be submitted in the prescribed forms. Digital signatures shall be used. Where there is conflict between the words and the figures, the words shall govern.
- b) All notations must be in ink or type written. No erasing or overwriting will be permitted. Mistakes may be crossed out and corrections typed or written with ink adjacent thereto and must be initialed in ink by the person or persons signing the tender.
- c) Tenders shall not contain any recapitulation of the work to be done. Alternative proposals will not be considered unless called for. No written, oral, telegraphic or telephonic proposals for modifications will be acceptable.
- d) Tenders shall be uploaded as notified on or before the date and time set for the opening of tenders in the Notice Inviting Tenders.
- e) Tenders subject to any conditions or stipulations imposed by the bidder are liable to be rejected.
- f) Each and every page of the tender document must be signed with company seal by the bidder.
- g) Any bidder may withdraw his tender by written request at any time prior to the scheduled closing time for receipt of tenders and not thereafter.

## **56. PRE-DISPATCH INSTRUCTION**

All instruments supplied against the purchase order shall be subjected to Inspection, check and /or test by the authorized representative from DHWU.

## **57. FINAL INSPECTION**

Final inspection will be carried by the authorized representative from DHWU.

## **58. SITE INSPECTION**

Bidder can inspect (at their own cost) the sites if required, for which they have to take necessary permission from DHWU in writing. DHWU will take at least four days to organize such permission.

## **59. ERASURES OR ALTERNATIONS**

The offers with overwriting and erasures may make the tender liable for rejection if each of such overwriting/erasures/manuscript ions is not only signed by the authorized signatory of the bidder. There should be no hand-written material, corrections or alterations in the offer. Technical details must be completely filled up. Correct technical information of the product being offered must be filled in. Filling up of the information using terms such as "OK", "accepted", "noted", "as given in brochure/manual" is not acceptable. The Customer may treat offers not adhering to these guidelines as unacceptable. The Customer may, at its discretion, waive any minor non-conformity or any minor irregularity in the offer. This shall be binding on all bidders and the Tender Committee reserves the right for such waivers.

## **60. COMPLIANCE WITH LAW**

The contractor hereto agrees that it shall comply with all applicable union, state and local laws, ordinances, regulations and codes in performing its obligations hereunder, including the procurement of licenses, permits certificates and payment of taxes where required. The contractor shall establish and maintain all proper records (particularly, but without limitation, accounting records) required by any law, code/practice of corporate policy applicable to it from time to time including records and returns as applicable under labor legislation.

## **61. CLARIFICATION OF BIDS**

During evaluation of the bids, the Customer/Tender Committee, at its discretion may ask the bidder for clarification of its bid. The request for the clarification and the response shall be in writing (fax/email) and no change in the substance of the bid shall seek offered or permitted.

## **62. DEEMED ACCEPTANCE**

Deliverables will be deemed to be fully and finally accepted by Diamond Harbour Women's University.

### 63. QUALITY CONTROL

- The contractor is obliged to work closely with DHWU and act within its authority and abide by directive issued by them on implementation activities.
- The contractor will abide by the safety measures and free DHWU and from all demands or responsibilities arising from accident/loss of life, the cause of which is due to their negligence. The bidder will pay all indemnities arising from such incidents and will not hold DHWU responsible.
- The contractor will treat as confidential all data and information about the system/items, obtained in the execution of its responsibilities in strict confidence and will not reveal such information to any party without the prior written approval of DHWU.
- DHWU reserves the right to inspect all phases of contractor's operation to ensure conformity to the specifications. DHWU shall have inspectors or other duly authorized representatives made known to the contractor, present during the progress of the work and such representatives shall have free access to the work at all times. The presence or absence of representatives of DHWU does not relieve the contractor of the responsibility for quality control in all phases.
- The Court of Diamond Harbour/Kolkata only will have the jurisdiction to deal with and decide any legal matters or dispute whatsoever arising out of the contract.

### 64. SITE NOT READY

The bidder shall not be in any manner liable for any delay arising out of DHWU inability to make the site ready within the stipulated period.

### 65. GENERAL TERMS

- a) All the pages of the bid document including documents submitted therein must be duly signed and stamped failing which the offer shall be liable to be rejected.
- b) All the documents to be submitted by the bidder along with their offer should be duly authenticated by the person signing the offer and if at any point of time during procurement process or subsequently it is detected that documents submitted are forged/tampered/manipulated in any way, the total responsibility lies with the bidder and DHWU reserves the full right to take action as may be deemed fit including rejection of the offer and such case is to be kept recorded for any future dealing with them.
- c) No Technical/Commercial clarification will be entertained after opening of the tender.
- d) Overwriting and erasures may make the tender liable for rejection if each of such overwriting/erasures/manuscript is not only signed by the authorized signatory of the bidder. All overwriting should be separately written and signed by the authorized signatory of the bidder.
- e) Quantity mentioned in the tender document is indicative only and orders shall be placed subject to actual requirement. DHWU reserve the right to increase or decrease the quantity specified in the tender.
- f) DHWU reserve the right to reject or accept or withdraw the tender in full or part as the case may be without assigning the reasons thereof. No dispute of any kind can be raised the right of buyer in any court of law or elsewhere.
- g) DHWU reserve the right to ask for clarification in the bid documents submitted by the bidder. Documents may be taken if decided by the committee.
- h) Supporting technical brochures/catalogues indicating each feature in respect of offered model and make must be submitted along with the offer, in absence of which the offer is liable to be ignored.
- i) No dispute by the bidders in regard to Technical/Commercial points will be entertained by DHWU and decision taken by the Tender Committee will be final.
- j) Discrepancy in the amount quoted by the bidder due to calculation mistake, the unit rate shall be regarded as firm and the totaling or carry in the amount quoted by the bidder shall be corrected accordingly.
- k) The price offers shall remain firm within the currency of contract and no escalation of price will be allowed.
- l) The acceptance of the tender will rest with the accepting authority who is not bound to accept the lowest or any tender and reserves the right to reject in part or in full any or all tender(s) received and to split up the work among participants without assigning any reason thereof.
- m) The customer/DHWU at its discretion may extend the deadline for the submission of Bids.
- n) The Court of Diamond Harbour/Kolkata only will have the jurisdiction to deal with and decide any legal matters or dispute whatsoever arising out of the contract.

### 65. OTHER TERMS AND CONDITION

1. The selected vendor has to supply, install, maintain, provide training and commissioning of the requisite hardware, software and peripherals at University.
2. The equipment/product must confirm to the specifications given and of desired quality. Consistency in quality shall be maintained for the entire lot of products offered. All the required quantity of items in schedule of requirement shall be of the same brand and model. Substitution of any internal components or subsystems of the product by similar items of different manufacturers will not be accepted for whatsoever reason.
3. All the equipments shall be supplied with the relevant interface cables and necessary standard accessories.
4. The selected vendor shall be responsible for the quality of supply.
5. Non-removable stickers have to be pasted on all kind of materials /items with details of company name, address. Contact detail, service centre/ help line number and warranty expiry date.
6. The indenter's right to inspect, test and, where necessary, reject the Goods after the Goods arrival at delivery Site shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed, prior to the Goods shipment.
7. All items should be covered a warranty period of minimum one year.

## SECTION – E

### BID FORM

(Bidders are requested to furnish the Bid Form in the Format given in this section, filling the entire Blank and to be submitted on Letter Head in original)

To  
DIAMOND HARBOUR WOMEN'S UNIVERSITY  
Sarisha, Diamond Harbour Road,  
South 24 Parganas, 743368

Sub: Delivery and Installation of Chemistry Laboratory instruments of Diamond Harbour Women's University.

Dear Sir,

1. We the undersigned bidder/(s), having read and examined in details the specifications and other documents of the subject tender no. E-NIT/CHEMLAB/22/DHWU/2019 dated 19.07.2019, do hereby propose to execute the job as per specification as set forth in your Bid documents.
2. The prices of all items stated in the bid are firm during the entire period of job irrespective of date of completion and not subject to any price adjusted as per in line with the bidding documents. All prices and other terms & conditions of this proposal are valid for a period of 180 (one hundred eighty) days from the date of opening of bid. We further declare that prices stated in our proposal are in accordance with your bidding.
3. We confirm that our bid prices include all other taxes and duties and levies applicable on bought out components, materials, equipments and other items and confirm that any such taxes, duties and levies additionally payable shall be to our account.
4. Earnest Money Deposit: We have enclosed EMD in the form of Demand draft for a sum of Rs. 30000/- (DD no. \_\_\_\_\_ dated \_\_\_\_\_ drawn on \_\_\_\_\_).
5. We declare that items shall be executed strictly in accordance with the specifications and documents irrespective of whatever has been stated to the contrary anywhere else in our proposal.
6. If this proposal is accepted by you, we agree to provide services and complete the entire work, in accordance with schedule indicated in the proposal. We fully understand that the work completion schedule stipulated in the proposal is the essence of the job, if awarded.
7. We further agree that if our proposal is accepted, we shall provide a Performance Bank Guarantee of the value equivalent to ten percent (10%) of the Order value as stipulated in Financial Bid (BOQ).
8. We agree that DHWU reserves the right to accept in full/part or reject any or all the bids received without any explanation to bidders and his decision on the subject will be final and binding on Bidder.

Dated, this .....day of .....2019

Thanking you, we remain,

Yours faithfully

.....  
Signature

.....  
Name in full

.....  
Designation

Signature & Authorized Verified by

.....  
Signature

.....  
Name in full

.....  
Designation

.....  
Company Stamp

## SECTION – F

### TECHNO COMMERCIAL EVALUATION & AWARDING OF CONTRACT

#### 1. EVALUATION PROCEDURE

- The Eligibility Criteria (Section - B) will be evaluated by Tender Committee and those qualify will be considered for further evaluation.
- The Tender Committee shall verify the Technical Specification (Technical Specification with Compliance Statement, Section – J) Deviation in specification shall not be allowed. Bidder qualified in Technical Specification shall be considered for further evaluation.
- After qualifying in Technical Specification, qualified bidders will only be considered for Financial Bid evaluation.

#### 2. FINAL EVALUATION

Financial Proposal of the bidders qualifying in the evaluation of Technical specification will be evaluated. The bidder who has qualified in the Technical Specification evaluation and returns with lowest quote (L1) in financial bid would normally be awarded the contract subject to Post Qualification.

#### 3. AWARDING OF CONTRACT

An affirmative Post Qualification determination will be prerequisite for award of the contract to the lowest quoted bidder. A negative determination will result in rejection of bidder's bid, in which event the DHWU will proceed to the next lowest evaluated bidder to make a similar determination of that bidder's capability to perform satisfactorily. The successful bidder (s) will have to give security deposit in the form of Performance Bank Guarantee.

#### 4. POST QUALIFICATION

The determination will evaluate the Bidder's financial, technical, design, integration, customization, production, management and support capabilities and will be based on an examination of the documentary evidence of the Bidder's qualification, as well as other information DHWU deems necessary and appropriate. This determination may include visits or interviews with the Bidder's client's reference in its bid, site inspection, and any other measures.

## SECTION – G

### GUIDANCE FOR E-TENDERING

Instructions / Guidelines for electronic submission of the tenders have been annexed for assisting the Bidders to participate in e-Tendering.

- **Registration of Bidder:**  
Any Bidder willing to take part in the process of e-Tendering will have to be enrolled & registered with the Government e-Procurement System through logging on to <https://wbtenders.gov.in>. The Bidder is to click on the link for e-Tendering site as given on the web portal.
- **Digital Signature Certificate (DSC):**  
Each Bidder is required to obtain a Class-II or Class-III Digital Signature Certificate (DSC) for submission of tenders from the approved service provider of the National Informatics Centre (NIC) on payment of requisite amount. Details are available at the Web Site stated above. DSC is given as a USB e-Token.
- The Bidder can search & download N.I.T. & BOQ electronically from computer once he logs on to the website mentioned above using the Digital Signature Certificate. This is the only mode of collection of Tender Documents.
- **Participation in more than one work:**  
A prospective bidder shall be allowed to participate in the job either in the capacity of individual or as a partner of a firm. If, found to be applied severally in a single job all the applications will be rejected.
- **Submission of Tenders:**  
Tenders are to be submitted through online to the website stated above in two folders at a time, one in Techno Commercial Proposal & the other is Financial Proposal before the prescribed date & time using the Digital Signature Certificate (DSC). The documents are to be uploaded virus scanned copy duly Digitally Signed. The documents will get encrypted (transformed into non readable formats)

The proposal should contain scanned copies of the following in two covers (folders).



**Techno Commercial Cover:**

Technical Document1 (scanned & join in pdf format then upload)

1. Copy of Demand Draft of Earnest Money Deposit (EMD)

Technical Document2 (scanned & join in pdf format then upload)

1. N I T Declaration duly stamped & signed in letter head, Section - Q
2. Bid Form as per format (Section – E)

Technical Compliance (scanned & joins in pdf format then upload)

1. Technical Specification With Compliance Statement (Section – I)
2. Manufacturer Authorisation Form (Section – M)

**Financial Cover:**

BOQ will be downloaded and same will be uploaded with quoted rates. While uploading BOQ file name shall remain unchanged. Absence of this document shall lead to summary rejection of the bid.

**NON-STATUTARY COVER (MY SPACE) CONTAIN FOLLOWING DOCUMENT:**

(In each folder, scanned coy will be uploaded with single file having multiple pages)

Sl. No.	Category Name	Sub Category Name	Sub Category Description
A	CERTIFICATES	A1. CERTIFICATES	<ul style="list-style-type: none"> <li>· GST Registration Certificate</li> <li>· PAN</li> <li>· Trade License</li> <li>· Document as per Clause no. 2 of Section – B</li> </ul>
B	COMPANY DETAILS	B1. COMPANY DETAILS 1	<ul style="list-style-type: none"> <li>· Document as per Clause – 1 of Section – B</li> </ul>
		B2. COMPANY DETAILS 2	<ul style="list-style-type: none"> <li>· Company Profile (Not more than 3 pages)</li> </ul>
C	CREDENTIAL	CREDENTIAL 1	Order copies as per Clause no. 4 of Section – B
		CREDENTIAL 2	<ul style="list-style-type: none"> <li>· Product brochure</li> <li>· Other documents, if any</li> </ul>
D	DECLARATION	DECLARATION 1	List of Clients as per format (Section – O)
		DECLARATION 2	Financial Capability of Bidder as per format (Section – K)
		DECLARATION 3	Bidder's Details as per format (Section – L)
		DECLARATION 4	Details of Order Executed as per format (Section – J)
F	FINANCIAL INFO	P/L & BALANCE SHEET 2015-2016	P/L & BALANCE SHEET 2015-2016
		P/L & BALANCE SHEET 2016-2017	P/L & BALANCE SHEET 2016-2017
		P/L & BALANCE SHEET 2017-2018	P/L & BALANCE SHEET 2017-2018

**SECTION-H**  
**BILL OF MATERIAL**

Sl. No.	Item Description	Qty	Unit
1	FT-IR Spectrometer with ATR	1	Nos.
2	UV-Visible Spectrometer	1	Nos.
3	Fluorescence Spectrophotometer	1	Nos.
4	Magnetic Susceptibility balance	1	Nos.
5	Rota evaporator with Chiller and Vacuum pump	2	Nos.
6	Low temperature reaction bath	1	Nos.
7	PC controlled Electrochemical Workstation	1	Nos.
8	Ultra low Freezer (-20 °C)	2	Nos.
9	Hot air oven	2	Nos.
10	Digital Balance (4 digit)	2	Nos.
11	Digital Balance (3 digit)	2	Nos.
12	Magnetic stirrer with hot plate.	10	Nos.
13	Muffle Furnace	1	Nos.
14	Ice Flaker machine (30kg):	1	Nos.
15	High Vacuum pump	2	Nos.
16	Vacuum pump	4	Nos.
17	Trinocular Stereozoom Microscope with LED transmitted & incident light application	1	Nos.

Bill of quantity may change at the time of ordering of Purchase Order. Detailed Technical Specifications are given in Section - I

**SECTION - I**

**TECHNICAL SPECIFICATION WITH COMPLIANCE STATEMENT**

(Tender No. E-NIT/CHEMLAB/22/DHWU/2019)

**1. Specification of Instruments:**

Sl. No.	Items	Specification details (Complete Exptal. Set)	Unit / Qty	Specification (Quoted / Applicable- by the bidder)	Complied (Yes / No)
1.	<b>FT-IR Spectrometer with ATR</b>	Full mid-IR wave number range from 7,800cm-1 to 350 cm-1 Michelson's interferometer with 300 incident angle, completely sealed and desiccated optics Signal-to-Noise ratio of minimum 30,000:1 or higher (peak-to-peak) Frictionless Flexible moving mirror mechanism for smooth moving mirror	1		

		<p>motion and high quality IR spectra  Maximum spectral resolution of 0.5 cm-1 selectable in steps of 0.5, 1, 2, 4, 8 &amp; 16 cm<sup>-1</sup>. Resolution must be achieved without any mask or aperture or apodised.  KBr beam splitter with moisture resistant coatings on surfaces  Must have high intensity long life Ceramic IR source  Should have high sensitivity temperature controlled DLATGS /DTGS detector  Data sampling should be done by stable He-Ne laser  Suitable maximum protection to beam splitter including suitable Dehumidifier or built-in Auto dryer for stable operation in extreme humidity conditions  Must have built-in validation software conforming to ASTM and European Pharmacopoeia.  System must be supplied with 32-bit Windows 2000 based FTIR software for complete instrument control full data processing including Arithmetic calculations; Dynamic spectral subtraction; Peak detection; Baseline correction; Normalization; Derivatives etc.  Must be supplied with advanced data processing capabilities include Full Quantitation; Spectrum search; Private library search; Multi-linear regression; Deconvolution; Kubelka-Munk conversion; Kramer-Kronig Analysis; ATR-correction; JCAMP conversion; ASCII conversion; Contaminant / Foreign Material Analysis Program and Pharma Report Program etc.  Should have built-in atmospheric correction function for automatic elimination of water vapor and CO2 peaks  Should have Automatic accessory recognition (ATR recognition) function with optional Quick-Start accessories  Should have Built-in self diagnostic function for checking, logging and status monitoring of interferometer conditions and key components in the interferometer &amp; sequential display of result  Must have Built-in security functions to meet the requirements of USB interface for PC connectivity  Diamond ATR must be offered as standard. Diamond crystal must be made of pure diamond  NaCl Demountable cell- 1 no must be supplied.  KBr die set, 15 Ton hydraulic press, Pallet holder, Agate Mortar &amp; Pestle etc. must be quoted by the vendor  10 cm Gas cell (NaCl) – 1 no. must be supplied</p> <p>USB cable, Branded PC with original OS,1 kVA online UPS must be offered along with Main system  Full system must be guaranteed for minimum period of 24 months from the date of installation.  Installation &amp; training must be provided to user upto their satisfaction at lab</p>			
2.	<b>UV-Visible Spectrometer</b>	<p>Technical Specification for UV-Visible Spectrometer with accessories</p> <p>Microprocessor based UV-VIS Spectrophotometer with high resolution touchscreendisplay, for operation on 220V / 50Hz.  Stand-alone operation or complete control through PC with PC software supplied asstandard  True double beam optics with aberration corrected concave blazed holographic gratingin Czerny – Turner mounting for high energy throughput and high qualitymonochromatic light  Wide wavelength range of 1,100 nm to</p>	1		

		<p>190 nm  High resolution 1 nm spectral bandwidth over entire wavelength range  Wavelength setting and display in steps of 0.1nm  Wavelength accuracy of + 0.1nm for D 2 spectral line  Wavelength reproducibility of + 0.1nm  Wavelength Slew rate: <math>\geq 14,500</math> nm/min  Variable wavelength scanning speed: <math>\geq 3,000</math> nm/min to 2 nm/min  29,000 nm/min when survey scanning  Ultra low stray light of <math>\leq 0.02\%T</math> at 220nm with NaI filter  Wide Photometric range of -4 to +4 Abs and 0 to 400 %T  High Photometric Accuracy of + 0.002 Abs at 0.5 Abs  High Photometric Repeatability of Less than <math>\pm 0.0002</math> Abs at 0.5 Abs  Baseline stability: <math>\leq 0.0003</math> Abs/Hr (700 nm, one hour after light source turned ON)  Ultra low Photometric noise of <math>\leq 0.00005</math> Abs (700 nm)  Dual source – high intensity Tungsten-Halogen and Deuterium lamp with automatic changeover  High sensitivity matched pair Silicon Photodiode detector  4 USB ports or more for high speed PC and printer connectivity, data storage and transfer through USB pen drive  Guaranteed compliance with all Pharmacopoeia requirements  Built in validation program, diagnostic and security functions  All operational modes as standard Photometric; Spectrum; Quantitation; Kinetics, Time Scan, DNA and Protein Quantitation in standalone and PC mode. Additionally Multi-Component measurement available in stand-alone mode.  Large sample compartment compatible with wide range of accessories.  Must supply one pair of 10mm path length Quartz Cuvettes of 3.5 ml volume as a standard supply  Branded PC, Printer and 1 KVA Online UPS</p>		
3.	<b>Fluorescence Spectrophotometer</b>	<p>Ideal for Fluorescence data &amp; Spectrum Analysis.  Stand alone operation  Ease of Use: Optional PC operation  Leading edge sensitivity (800 or better RMS)  An Excellent signal-to-noise ratio (800 S/N RMS, Raman band of water at 5 nm bandwidth).  PC control the scan speed is as high as 12,000nm/min, making it ideal for 3-D scan measurements.  A wide array of accessories is available, including automatic sampling, quantum yield measurement, full-range spectral correction, and 100uL micro cell.  Wavelength range : 220-730 nm (zero order light, EX and EM)  Sensitivity : <math>S/N \geq 800</math> (RMS) Raman band of water, Ex. 350nm,  Slit width 5nm, Response 2s  Background sensitivity : <math>S/N</math> more than 10000 :1  Wavelength Accuracy : <math>\pm 2.0</math> nm  Band pass : 2.5, 5.0, 10, 20 nm (EX and EM)  Wavelength Scan Speed : 60, 300, 1500, 3000 and 12000 nm/min (12,000 nm/min only under PC control)  Wavelength repeatability : <math>\leq 1</math> nm  Wavelength drive speed : 12,000</p>	01	

		<p>nm/min Sensitivity : 800 S/N RMS Photometric principal : Monochromatic light monitoring ration calculation Light Source : 150W Xenon lamp Power consumption : AC100, 115, 220, 230, 240V 50/60Hz 400VA Indigenous Items Branded PC, Printer and 2 KVA Online UPS</p>			
4.	<b>Magnetic Susceptibility balance</b>	<p>Technical Specification for Magnetic Susceptibility balance with accessories</p> <p>Measurement ranges <math>\chi_v</math> and <math>\chi_g</math> in 4 ranges from 0 to <math>\pm 1.999 \times 10^{-7}</math>, <math>10^{-6}</math>, <math>10^{-5}</math>, <math>10^{-4}</math> c.g.s. units Sample volume required - 10.200 <math>\mu</math>l Magnetic field strength - 4.5 kGauss Analogue output - Voltage proportional to display value via BNC connector on detector unit Back-panel (<math>\pm 3V</math>). Main Unit, Calibration Standard Tube, 2 Normal Sample Tubes and Operating Manual Multi point PSU(220VAC 50Hz), Software latest version</p>	01		
5.	<b>Rota evaporator with Chiller and Vacuum pump</b>	<p><b><u>A. Vertical Condenser Rotary Vacuum Evaporator:</u></b> Speed range : 10 to 310 rpm Provision for clock wise and anti clock wise motion as available. Evaporation capacity : Max. 23ml/min(Water evaporation) Rotation speed setting and display : Setting by dial Digital display Jack function : Manual jack function(Jack stroke 180mm, stepless) Motor : DC Brush less motor Vacuum seal : Vacuum seal(Teflon® +Teflon®. Viton double seal 1set. Sample Flask : 1 L Pear Style Ts 29/38 Receiving Flash: 1 L round style Condenser: Vertical double spiral glass coil</p> <p><b><u>B. It should comes complete with Digital Water Bath</u></b> Temp. control range : Ambient temp. 5°C to 90°C Bath Inlet terminal : For connection to Evaporator main unit driven part Max 2A Bath dimension : <math>\varnothing</math> 220 x 120H, Capacity : 4.3 L</p> <p><b><u>C. Powerful Aspirator Vacuum Pump</u></b> Displacement method : Pressured water jets Displacement: 16L/min. x 2 (at water temperature 6°C) Ultimate vacuum : Max. Vacuum 7mm of mercury Motor: Induction motor output 150W Water bath capacity : 10L</p> <p><b><u>D. COMPACT LOW TEMPERATURE CIRCULATOR</u></b> Temp. Control range : -20°C to Ambient Temp. Control and display : Digital Refrigeration unit : 450W Bath capacity : 4L Circulating pump : Max. Flow 10L/min., Pressure Head 4.2m Foot Print : 230W x 470D x 490H mm</p>	02		
6.	<b>Low temperature reaction bath</b>	<p>Temp. control range : -80°C to 0° C Temp. control : Heater, P.I.D. control, Refrigerator, Continuous operation Magnetic Stirrer : 450W Temp. setting, display : Digital display, Membrane key switch, Safety : Leak/over current breaker, Independent overheat protector, Self diagnosis function, Refrigerator protection timer, Overload relay protection circuit,</p>	01		

		<p>Float Switch.  Magnetic Stirring/Rotation range: 100 ~800rpm  Independent Digital Rotation Setting &amp; Monitoring: Digital  Operation Time : Auto start &amp; stop timed  Refrigerator. : Air cooling 450W x 2  <b>Cooling capacity</b> : 400W @ -80°C  Bath dimension : ø180 x 140D mm, 3.5L  Range of vessel size : 50 - 300ml  <b>Its comes complete with lid w/variable opening shutter, clamp sets for standard flasks.</b>  To set/change any temperature range and rotation speed of Magnetic Stirrer at any instant of time without opening the bath cover and without disturbing the reaction process/parameters.  Compact bench top design to be used inside of Fume Hood also.</p> <p><b>Footprint Dimensions</b> : 20-22"W x 16-20" D</p>			
7.	<b>PC controlled Electrochemical Workstation</b>	<p>Technical Specification for PC controlled Electrochemical Workstation with accessories  <b>Hardware :</b>  Output Current : 1000 mA  Current Resolution : 1 nano Amps.  Polarisation Potential : 5 Volt  Output Potential (Compliance ) : 20 Volt  Maximum Scan Rate :3000 mV / sec  Minimum Scan Rate : 0.1mV / sec  Supply : 220 – 240 Volt, 50 Hz  <b>Software :</b> Windows based software with following measurement techniques  Measurement of rest potential  Potentiostatic Potentiodynamic  Linear Polarisation Tafel Plot Corrosion Calculator  Cyclic Voltammety  Galvanostatic Coulometry  Dynamic Capacitance  Solar cell Efficiency  Charging and discharging of Cell  <b>Accessories :</b>  (a) Corrosion Cell with Platinum and SCE Electrode  (b) Branded Intel Dual Core Personal Computer with 2GB RAM, 500GB HDD, Sound card, speaker, DVD writer, 18.5" LED Colour Monitor, Key Board and Optical Mouse.  (c) Nitrogen gas cylinder with regulator</p>	01		
8.	<b>Ultra low Freezer (-20 °C)</b>	<p>Technical Specification for Ultra low Freezer (-20 °C)with accessories  Capacity (in litres) : 142  Dimension HxWxD (in inches) :45x21x22  Temp. Range : -18°C ~ -22°C  Wheels : No  Lock : No  Defrost : Manual  Rated Load, Wattage : 125  No. of Shelves / Drawers : 5  No. of Lids/Doors : Complete with Digital display &amp; 1 kvstabiliser</p>	02		
9.	<b>Hot air oven</b>	<p>Technical Specification for Hot air oven with accessories  Inner Chamber: Stainless Steel of 304 Grade with Mirror Finish.  Inner Chamber Size: 18" x 18" x 24  Temp. Range 5 0C above Ambient to 200 C.  Temperature Microprocessor based PID  Temperature indicator cum controller,  LED Display of Set Value (SV)  Control Process Value (PV), with auto</p>	02		

		tune Heating Elements: Heating Element Kanthal -1 are place at the bottom & sidewall of the oven chamber. Door Front opening door is fitted with Polish Head Proof Handle with latch locking arrangement Number of Shelves 3 Nos. of removable & Adjustable Selves made ofStainless Steel Wire Mesh.			
10.	<b>Digital Balance (4 digit)</b>	Technical Specification forDigital balance with accessories Maximum capacity 220 g Weighing platform dimensions 90 mm Readability 0.1 mg Repeatability 0.1 mg Linearity 0.2 mg Settling time 2s Sensitivity temperature drifts 2.0 ppm/°C Weight of balance 4.7 kg (ME) / 4.5 kg (ME-E) Selection of non-critical materials –PVC-free, mercury-free Overload protection up to 100kg Solid metal base Reinforced body 10 built-in applications	02		
11	<b>Digital Balance (3 digit)</b>	Technical Specification for Digital balance with accessories Maximum capacity 220g Weighing platform dimensions 90 mm Readability 1 mg Repeatability 0.1 mg Linearity 0.2 mg Settling time 2 s Sensitivity temperature drift2.0 ppm/°C Weight of balance 4.7 kg (ME) / 4.5 kg (ME-E) Selection of non-critical materials –PVC-free, mercury-free Overload protection up to 100kg Solid metal base Reinforced body 10 built-in applications	02		
12	<b>Magnetic stirrer with hot plate.</b>	Technical Specification for Magnetic stirrer with hot plate accessories Capacity: 5 Litres with Hot Plate. With Stepless Speed control & S.S.body and top. Max. Speed : 1200 RPM.	10		
13	<b>Muffle Furnace</b>	Technical Specification for <b>Muffle Furnace</b> with accessories  Electrically heated rectangular shape horizontally chamber type Muffle furnace shall be made for heat treatment and Laboratory testing purpose. Temperature shall be uniform throughout the chamber in between 5 °C above ambient .Suitable temperature up to 1000°C Maximum and 950°C working condition. The effective chamber size (cavity) shall be made6” width 6” height and 12” depth horizontally type. <b>Temperature:-</b> The design of the furnace temperature 1000°C maximum & working at 950 °C continuous. <b>Control System:</b> - A Digital controller shall be provided with the furnace. It will be control the temperature at any setting in between Room temperature at 1000 °C automatically. <b>Control accuracy ± 1 °C.</b> <b>Power consumption:</b> The power consumption of the furnace 3.0 K.W. Load approximate, 220/230 Volt, single phase with Neutral Ac supply, 50 Hz. <b>Heating element:</b> The heating element	01		

		<p>(coil) shall be made by the <b>Imported Sweden make Kanthal A-1 wire</b> of 16 S.W.G for long life. The coil shall be provided in the Iron less outside the muffle. All the connection of the coil shall be mounted on the non-contractor board at backside fore a maintenance.</p> <p><b>Insulation details:</b> The hot face insulation shall be lined in the chamber and lightweight cold face insulation shall be line with the first layer with the body. The ceramic fiberboard shall be provided in the between the insulation lining. The skin temperature should not be excessive of 70 °C as per I.S. Standard.</p> <p><b>Door system:</b> The door of the furnace shall be provided at front. The manual operated door shall be open at a side. The hinges type airtight door shall be provided with the furnace. A peephole shall be provided in middle for observation. Control panel box and electrical accessories: The control panel built shall be made with the furnace bottom with hinges door shall be provided at front. The panel shall be complete with the following items as per good engineering practices.</p> <p><b>CONTROL PANEL:-</b> The control panel unit shall be made with the furnace bottom of the chamber complete with the following items as per good engineering practice.</p> <p><b>1 No. Digital Temperature Indicator cum Controller.</b> 1 No. Line indicating lamps for indicate the incoming line. 1No. Protective Fuse</p> <p><b>Painting.:</b> All the Mild steel part shall be painted by the two coats of rush protective red oxide primer and finished by the two coats of anticorrosive-epoxy spray painting.</p>			
14	<b>Ice Flaker machine (30kg):</b>	Capacity - 30 kg. Storage Capacity - 15 kg. Power Supply - AC 220V / 50Hz Dimension - 300x543x722 mm.	01		
15	<b>High Vacuum pump</b>	Two stage oil sealed, rotary vane vacuum pump with 0.55kW single phase motor, 220V±10%, 50 Hz, IP44 Ult. No load Vacuum.: 0.001 mm of Hg <b>With</b> Digital pirani gauge with single gauge head, 2m cable - DHPG-222 And SS Bellow with KF25 end flanges - Ø 1" x 1000mm L	02		
16	<b>Vacuum pump</b>	Complete with electric motor, Air ballast, first filling Vacuum oil, Stage: Double Capacity: 100 lit/min. Motor Capacity : 1/4 H.P. Ult. No load Vacuum.: 0.001 mm of Hg (McLeod Gauge)	04		
17	<b>Trinocular Stereozoom Microscope with LED transmitted &amp; incident light application</b>	Technical Specification:  Model: "Stemi 305"  <ul style="list-style-type: none"> <li>● Zoom Stereo Microscope Body with zoom ratio 5:1 with zoom setting for 5 position (0.8x-1x-2x-3x-4x)</li> <li>● Variable magnification range between 8x to 40x built in 1x objective &amp; extended higher magnification by using additional auxiliary objectives &amp; Wide field focusable eyepieces 10x with 23 mm field of view &amp; inter-pupliary adjustment 55-75 mm.</li> <li>● Stand fitted with with broad base easy - to – reach controls for focusing, zoom changer and illumination Photoport with 100% switch between right eyepath and camera for 3D viewing, incl. interchangeable 60N camera adapter 0.5x</li> </ul>	01		



		<p>with c-mount interface</p> <ul style="list-style-type: none"> <li>• Ergonomic Binocular phototube inclined at 35° with built-in Plan-apochromatic objective &amp; eyepieces that cover large area of sample 35.4mm to 4.6 mm. Greenough based optical system of resolution 450 lp/mm</li> <li>• Working distance 110 mm.</li> <li>• Stand with Integrated LED trans-illumination base with working area 160x 195 mm &amp; column for bright field and . Also complete with spot illumination as well as oblique vertical LED illumination for focusable reflected light application. Lamp life of LED 25000 Hrs&amp; color temperature 5600K.</li> <li>• Optics are anti-fungus treated &amp; certified up to three years from the principal (certificate enclosed)</li> <li>• Adapter for attaching camera on camera port 60N</li> </ul>			
--	--	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	--	--	--

Authorized Signatory (Signature In full): \_\_\_\_\_

Name and title of Signatory: \_\_\_\_\_

Stamp of the Company: \_\_\_\_\_

**SECTION – J**

**DETAILS OF ORDERS EXECUTED BY BIDDER**

(Tender No. E-NIT/CHEMLAB/22/DHWU/2019)

Sl. No.	Order No.	Order Date	Order Value	Brief description of items and job details	Completed (Yes/NO)	Name of the Customer	Contact details of the Customer

Authorized Signatory (Signature In full): \_\_\_\_\_

Name and title of Signatory: \_\_\_\_\_

Stamp of the Company: \_\_\_\_\_

**Note:**

- A. Type of Project shall indicate the implementation of services.
- B. Scope of work shall indicate whether it is implementation, Operation or maintenance.
- C. Submit Customer Order Copy details of the order indicating the project value, customer contact details.

**SECTION – K**

**FINANCIAL CAPABILITY OF BIDDER**

(Tender No. E-NIT/CHEMLAB/22/DHWU/2019)

**FINANCIAL INFORMATION**

Sl. No.	Name of the Bidder	Turnover (Rs. / Crores)		
		2015-16	2016-17	2017-18
1				

Authorized Signatory (Signature In full): \_\_\_\_\_

Name and title of Signatory: \_\_\_\_\_

Stamp of the Company: \_\_\_\_\_

**Note:**

Submit the audited financial statement/ audited annual report of the last three financial years.

## SECTION – L

### BIDDERS'S DETAILS

(Tender No. E-NIT/CHEMLAB/22/DHWU/2019)

1	Name of the Firm	
2.	Registered Office Address	
	Contact Number	
	Fax Number	
	E-mail	
3.	Correspondence / Contact address	
	Name & Designation of Contact person	
	Address	
	Contact Number	
	Fax Number	
	E-mail	
4	Is the firm a registered company? If yes, submit documentary proof	
5	Year and Place of the establishment of the company	
6	Former name of the company, if any	
7	Is the firm <ul style="list-style-type: none"><li>• a Government / Public Sector Undertaking</li><li>• a propriety firm</li><li>• a partnership firm (if yes, give partnership deed)</li><li>• a limited company or limited corporation</li><li>• a member of a group of companies, (if yes, give name and address and description of other companies)</li><li>• a subsidiary of a large corporation (if yes give the name and address of the parent organization). If the company is subsidiary, state what involvement if any, will the parent company have in the project.</li></ul>	
8.	Is the firm registered with Sales Tax department? If yes, submit valid GST Registration certificate.	
9	Is the firm registered for Service Tax with Central Excise Department (Service Tax Cell)? If yes, Submit valid Service Tax registration certificate.	
10	Total number of employees. Attach the organizational chart showing the structure of the organization	
11	Are you registered with any Government/ Department/ Public Sector Undertaking (if yes, give details)	
12	How many years has your organization been in business under your present name?	
13	What type best describes your firm? (Purchaser reserves the right to verify the claims if necessary) <ul style="list-style-type: none"><li>• Manufacturer</li><li>• Supplier</li><li>• System Integrator</li><li>• Consultant</li><li>• Service Provider (Pl. specify details)</li><li>• Software Development</li><li>• Total Solution provider (Design, Supply, Integration, O&amp;M)</li><li>• IT Company</li></ul>	
14	Name of Offices in district head quarters in West Bengal	
15	Is your organization has ISO 9001:2008 certificates?	
16	List the major clients with whom your organization has been / is currently associated.	
17	Have you in any capacity not completed any work awarded to you? (If so, give the name of project and reason for not completing the work)	
18	Have you ever been denied tendering facilities by any Government / University / Public sector Undertaking? (Give	

	details)	
--	----------	--

Authorized Signatory (Signature In full): \_\_\_\_\_

Name and title of Signatory: \_\_\_\_\_

Company Rubber Stamp: \_\_\_\_\_

**SECTION – M**

**MANUFACTURER'S AUTHORIZATION FORM**

Date:

To  
The Registrar  
DIAMOND HARBOUR WOMEN'S UNIVERSITY  
Sarisha, Diamond Harbour Road,  
South 24 Parganas, 743368

Ref: Tender No.: E-NIT/CHEMLAB/22/DHWU/2019 dated 19.07.2019

WHEREAS \_\_\_\_\_ who are official producers of  
\_\_\_\_\_ and having production facilities at  
\_\_\_\_\_ do hereby authorize  
\_\_\_\_\_ located at  
\_\_\_\_\_ (hereinafter, the "Bidder") to submit a bid of the following Products  
produced by us, for the Supply Requirements associated with the above Invitation for Bids.

When resold by \_\_\_\_\_, these products are subject to our applicable standard end user warranty terms.

We assure you that in the event of \_\_\_\_\_, not being able to fulfill its obligation as our Service Provider in respect of our standard Warranty Terms we would continue to meet our Warranty Terms through alternate arrangements.

We also confirm that \_\_\_\_\_ is our authorized service provider and can hence provide maintenance and upgrade support for our products.

We also confirm that the products quoted are on our current product list and are not likely to be discontinued within 3 years from the day of this letter. We assure availability of spares for the products for the next two years after three years warranty.

We also confirm that the material will be delivered as mentioned from the date of placement of confirmed order.

Name \_\_\_\_\_ In the capacity of \_\_\_\_\_

Signed \_\_\_\_\_

Duly authorized to sign the authorization for and on behalf of \_\_\_\_\_

Dated on \_\_\_\_\_ day of \_\_\_\_\_ 2019

Note: This letter of authority must be on the letterhead of the Manufacturer and duly signed by an authorized signatory.

**SECTION – N**

**FORMAT FOR PRE-BID MEETING QUERY**

(Tender No. E-NIT/CHEMLAB/22/DHWU/2019)

**This is deleted from the tender document and is not required to submit.**

**SECTION – O**

**LIST OF CLIENTS OF SIMILAR ORDERS**

(Tender No. E-NIT/CHEMLAB/22/DHWU/2019)

Sl. No.	Name of the Client	Address	Contact Person	Designation	Contact Numbers

Authorized Signatory (Signature In full): \_\_\_\_\_

Name and title of Signatory: \_\_\_\_\_

Company Rubber Stamp: \_\_\_\_\_

**SECTION – P**

**PROFORMA FOR PERFORMANCE BANK GUARANTEE**

(On non-judicial stamp paper of appropriate value to be purchased in the name of executing Bank)

PROFORMA OF BANK GUARANTEE FOR SECURITY DEPOSIT –CUM-PRFORMANCE  
GUARANTEE

Ref ..... Bank Guarantee no.....

Date.....

PROFORMA OF BG FOR SECURITY DEPOSIT

KNOW ALL MEN BY THESE PRESENTS that in consideration of DIAMOND HARBOUR WOMEN’S UNIVERSITY, a Government of West Bengal Undertaking incorporated under the DHWU Act, 2012 having its Registered office at Sarisha, Diamond Harbour Road, 24 Pgs (S), 743368 (hereinafter called “The Purchaser”) having agreed to accept from \_\_\_\_\_(hereinafter called “The Contractor”) Having its Head Office at \_\_\_\_\_, a Bank guarantee for Rs. \_\_\_\_\_ in lieu of Cash Security Deposit for the due fulfillment by the Contractor of the terms & conditions of the Work Order No.

\_\_\_\_\_ dated \_\_\_\_\_ issued by the Purchaser for \_\_\_\_\_ (hereinafter called “the said work order \_\_\_\_\_ dated \_\_\_\_\_)”. We \_\_\_\_\_ (Name & detailed address of the branch) (hereinafter called “the Guarantor” ) do hereby undertake to indemnify and keep indemnified the Purchaser to the extent of Rs. \_\_\_\_\_ (Rupees \_\_\_\_\_) only against any loss or damage caused to or suffered by the Purchaser by reason of any breach by the Contractor of any of the terms and conditions contained in the said Work Order No. \_\_\_\_\_ dated \_\_\_\_\_ of which breach the opinion of the Purchaser shall be final and conclusive.

(2) AND WE, \_\_\_\_\_ DO HEREBY Guarantee and undertake to pay forthwith on demand to the Purchaser such sum not exceeding the said sum of \_\_\_\_\_ Rupees \_\_\_\_\_) only as may be specified in such demand, in the event of the Contractor failing or neglecting to execute fully efficiently and satisfactorily the order for \_\_\_\_\_ Work Order no. , \_\_\_\_\_ dated \_\_\_\_\_

(3) WE \_\_\_\_\_ further agree that the guarantee herein contained shall remain in full force and effect during the period that would be taken for the performance of the said order as laid down in the said Work Order No. \_\_\_\_\_ dated \_\_\_\_\_ including the warranty obligations and that it shall continue to be enforceable till all the dues of the Purchaser under or by virtue of the said Work Order No. \_\_\_\_\_ dated \_\_\_\_\_ have been fully paid and its claims satisfied or is charged or till the Purchaser or its authorized representative certified that the terms and conditions of the said Work Order No. \_\_\_\_\_ dated \_\_\_\_\_ have been fully and properly carried out by the said contractor and accordingly discharged the guarantee.

(4) We \_\_\_\_\_ the Guarantor undertake to extend the validity of Bank Guarantee at the request of the contractor for further period of periods from time to time beyond its present validity period failing which we shall pay the Purchaser the amount of Guarantee.

(5) The liability under the Guarantee is restricted to Rs. \_\_\_\_\_ (Rupees \_\_\_\_\_) only and will expire on \_\_\_\_\_ and unless a claim in writing is presented to us or an action or suit to enforce the claim is filled against us within 6 months from \_\_\_\_\_ all your rights will be forfeited and we shall be relieved of and discharged from all our liabilities (thereinafter)

(6) The Guarantee herein contained shall not be determined or affected by liquidation or winding up or insolvency or closer of the Contractor.

(7) The executants has the power to issue this guarantee on behalf of Guarantor and holds full and valid power of Attorney granted in his favour by the Guarantor authorizing him to execute the Guarantee.

(8) Notwithstanding anything contained herein above, our liability under this guarantee is restricted to Rs. \_\_\_\_\_ (Rupees \_\_\_\_\_) only and our guarantee shall remain in force up to \_\_\_\_\_ and unless a demand or claim under the guarantee is made on us in writing on or before \_\_\_\_\_ all your rights under the guarantee shall be forfeited and we shall be relieved and discharged from all liabilities there under.



WE, \_\_\_\_\_ lastly undertake not to revoke this guarantee during the currency except with the previous consent of the Purchaser in writing. In witness whereof we \_\_\_\_\_ have set and subscribed our hand on this \_\_\_\_\_ day of \_\_\_\_\_.

SIGNED, SEALED AND DELIVERED

\_\_\_\_\_  
(Stamp of the executants)

WITNESS

1) \_\_\_\_\_

2) \_\_\_\_\_  
(Name & address in full with Rubber Stamp)

### **INSTRUCTIONS FOR FURNISHING BANK GUARANTEE**

1. Bank Guarantee (B.G.) for Advance payment, Mobilization Advance, B.G. for security Deposit-cum-Performance Guarantee, Earnest Money should be executed on the Non- Judicial Stamp paper of the applicable value and to be purchased in the name of the Bank.
2. The Executor (Bank authorities) may mention the Power of Attorney No. and date of execution in his/her favour with authorization to sign the documents. The Power of Attorney is to be witnessed by two persons mentioning their full name and address.
3. The B.G. should be executed by a Nationalised Bank/ Scheduled Commercial Bank preferably on a branch located in Sarisha. B.G. from Co-operative Bank / Rural Banks is not acceptable.
4. A Confirmation Letter of the concerned Bank must be furnished as a proof of genuineness of the Guarantee issued by them.
5. Any B.G. if executed on Non-Judicial Stamp paper after 6 (six) months of the purchase of such stamp shall be treated as Non-valid.
6. Each page of the B.G. must bear signature and seal of the Bank and B.G. Number.
7. The content of the B.G. shall be strictly as Proforma prescribed by DHWU in line with Purchase Order /LOI/ Work Order etc. and must contain all factual details.
8. Any correction, deletion etc. in the B.G. should be authenticated by the Bank Officials signing the B.G.
9. In case of extension of a Contract the validity of the B.G. must be extended accordingly.
10. B.G. must be furnished within the stipulated period as mentioned in Purchase Order / LOI / Work Order etc.
11. Issuing Bank / The Bidder are requested to mention the Purchase Order / Contract / Work Order reference along with the B.G. No. For making any future queries to DHWU.

**SECTION - Q**

**NIT DECLARATION**

(Bidders are requested to furnish the Format given in this section, filling the entire Blank and to be

submitted on Bidder's Letter Head)

To  
The Registrar  
DIAMOND HARBOUR WOMEN'S UNIVERSITY  
Sarisha, Diamond Harbour Road,  
South 24 Parganas, 743368

Sub: Delivery and Installation of Chemistry Laboratory instruments of Diamond Harbour  
Women's University.

Dear Sir,

We the undersigned bidder/(s) declare that we have read and examined in details the specifications and other documents of the subject tender no. E-NIT/CHEMLAB/22/DHWU/2019 dated 19.07.2019 Delivery and Installation of Chemistry Laboratory instruments published by DIAMOND HARBOUR WOMEN'S UNIVERSITY in e-Tender website.

We further declare that we have agreed and accepted all the clauses / sub-clauses / formats / terms & conditions other requirements related to the said tender and we will abide by the same as mentioned in the tender document while participating and executing the said tender.

Thanking you, we remain

Yours faithfully

.....  
Signature

.....  
Name in full

.....  
Designation

.....  
Company Stamp

Dated, this .....day of .....2019